



HAL
open science

Atomic layer deposition for biosensing applications

Octavio Graniel, Matthieu Weber, Sebastien Balme, Philippe Miele, Mikhael Bechelany

► **To cite this version:**

Octavio Graniel, Matthieu Weber, Sebastien Balme, Philippe Miele, Mikhael Bechelany. Atomic layer deposition for biosensing applications. *Biosensors and Bioelectronics*, 2018, 122, pp.147 - 159. 10.1016/j.bios.2018.09.038 . hal-01919095

HAL Id: hal-01919095

<https://hal.umontpellier.fr/hal-01919095>

Submitted on 4 Jun 2021

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Atomic Layer Deposition for Biosensing Applications

Octavio Graniel¹, Matthieu Weber¹, Sébastien Balme¹, Philippe Miele^{1,2} and Mikhael Bechelany^{*1}

¹Institut Européen des Membranes IEM, UMR-5635, Université de Montpellier, ENSCM, CNRS, Place Eugène Bataillon, F-34095 Montpellier Cedex 5, France

²Institut Universitaire de France (IUF), MESRI, 1 rue Descartes, 75231 Paris cedex 05, France

* Corresponding author: mikhael.bechelany@umontpellier.fr

Abstract

Atomic layer deposition (ALD) is a thin film deposition technique currently used in various nanofabrication processes for microelectronic applications. The ability to coat high aspect ratio structures with a wide range of materials, the excellent conformality, and the exquisite thickness control have made ALD an essential tool for the fabrication of many devices, including biosensors. This mini-review aims to provide a summary of the different ways ALD has been used to prepare biosensor devices. The materials that have been deposited by ALD, the use of the ALD layers prepared and the different types of biosensors fabricated are presented. A selected list of studies will be used to illustrate how the ALD route can be implemented to improve the operational performance of biosensors. This work comprehensively shows the benefits of ALD and its application in various facets of biosensing and will help in exploiting the numerous prospects of this emerging and growing field.

1. Introduction

Atomic layer deposition (ALD) is a vapor deposition technique enabling the preparation of thin film materials with high conformality and excellent control over the thickness (Chalker, 2016; George and Steven M. George, 2010; Ritala and Leskelä, 2002). Its origins can be traced back to two different places (Puurunen, 2014). ALD was first developed by Aleskovskii during the 1960s in Russia and was referred to as molecular layering (ML) (Malygin et al., 2015). Later on, at the beginning of the 1970s in Finland, Suntola developed the atomic layer epitaxy (ALE) technique to deposit ZnS for electroluminescent displays (Puurunen, 2014; Suntola and Antson, 1977). In the next decades and up to today, the route was then referred to as ALD.

ALD is based on self-limiting reactions between two gaseous precursors and allows the deposition of thin films in a layer-by-layer fashion. The ability to deposit conformal films on high aspect ratio structures, with high uniformity over large areas, at (relatively) low temperatures, has made ALD a technique of choice for the preparation of ultrathin films and a key enabling technology (George and Steven M. George, 2010; Leskelä and Ritala, 2003; Ritala and Leskelä, 2002). ALD allows the deposition of a wide range of materials such as oxides (George and Steven M. George, 2010; Hämäläinen et al., 2014), nitrides (Kim, 2003; Weber et al., 2017), sulfides (Meng et al., 2017; Peters et al., 2015), and pure elements (Aaltonen et al., 2004; Johnson et al., 2014; Lim et al., 2003; Weber et al., 2014). These features

have made ALD a relevant technique for many applications such as fuel cells (Gong et al., 2013), metal oxide semiconductor field effect transistors (MOSFETs) (Hong et al., 2016), water splitting (Ho et al., 2018) and purification (Weber et al., 2018), encapsulation (Black et al., 2018), membranes (Weber et al., 2017), solar cells (Elias et al., 2012; Van Delft et al., 2012), and batteries (Liu et al., 2014). Also, the possibility of controlling the composition of the deposited layer by making nanolaminate or alloyed structures allows the tailoring of the chemical and physical properties of the ALD prepared materials (Chaaya et al., 2014; Gu et al., 2013).

ALD is a cycle based process involving four steps (depicted schematically in Figure 1). In the first step, a precursor is introduced into the reactor chamber and is left enough time to react with the surface groups of the substrate. Next, the unreacted precursor molecules and the by-products are removed by purging/pumping the system with an inert gas (usually N₂ or Ar). The third step involves the introduction of a second precursor (the co-reactant) to react with the adsorbed molecules. The final step consists in purging/pumping again the system to remove unreacted precursor and by-products molecules. As a result, one (sub) monolayer of the desired material is deposited on the substrate surface. This cycle is then repeated until the desired thickness is obtained.




Figure 1. Schematic representation of the ALD process

The use of ALD in the fabrication of nanomaterials for biological and medical applications has proliferated and is now widely spread (Guo et al., 2010; Marichy et al., 2012; Pessoa et al., 2017; Vähä-Nissi et al., 2014; K. Zhang et al., 2017). The compatibility of ALD with the nanoscale of the components found in biomedical devices, the biocompatibility of the materials that can be deposited, and the tuning of the chemical reactivity are some of the reasons why there is a keen interest for using ALD for biomedical applications. Among these applications, biosensing has recently benefited from ALD as a tool for the fabrication of biosensors. Figure 2 shows a histogram of the number of publications (using the Web of Science platform) referring to “biosens*” and either “atomic layer deposition” or “ALD”. The survey clearly shows that the number of works relating both terms has been steadily increasing in the last decade.




Figure 2. Histogram showing the number of publications referring to “biosens*” and either “atomic layer deposition” or “ALD” using the Web of Science platform.

Generally speaking, biosensors are devices that allow the selective detection of a target molecule (Turner, 2013), making them useful for many applications such as clinical diagnosis (Fu et al., 2016), food safety (Luo et al., 2009), environmental monitoring (Mishra et al., 2017), security and bioterrorism (Liu et al., 2018; Muhammad-Tahir and Alocilja, 2003). A typical biosensor configuration is presented in Figure 3. Biosensors have two main components: a biorecognition layer and a physical transducer. When the analyte of interest (e.g., low molecular compound, (bio) macromolecules, protein, virus, cell) is captured by the biorecognition element, a biochemical signal is produced, and the transducer element transforms it into a signal that can be measured and correlated (in some cases) to the concentration of it. Biosensors are selective thanks to the bioselective layer that interacts only with the analyte of interest. They can be classified either by the type of biorecognition element (e.g., enzyme, antibodies, cell, nucleic acids) or by the nature of transducer signal (i.e., electrochemical, optical, electrical, mass-sensitive, magnetic).




Figure 3. Typical biosensor configuration

In the field of biosensing, ALD has been used mainly as a fabrication tool, or as a way to enhance the biosensing performance. In the next sections, a review of the different applications of ALD for these purposes is given. The type of materials deposited by ALD, the use of the ALD layers prepared, the different types of biosensors fabricated, and the detection methods used for sensing are reported.

2. ALD for biosensing applications

ALD has proven to be a versatile technique to fabricate nanostructures (Hong and Kim, 2016; Li et al., 2017; Viter et al., 2015; Xiong et al., 2015). The precise thickness control and excellent conformality allowed by the technique enables to fabricate complex nanostructures with a high surface area that can be used for the design of biosensors. Some of the films deposited by ALD can function as sacrificial layers that can be removed later by chemical or physical etching to produce the desired nanostructure. Other films can be used to fabricate the transducer element of a biosensor, which requires biocompatible materials that can interact with the bio-recognition elements. Additionally, they can be used to protect biosensors and improve their stability towards harsh environments. In this section, ALD films deposited by ALD used in combination with other techniques for nanofabrication will be discussed.

2.1 ALD to produce nanostructured biosensors

Recently, the interest in fabricating biosensors with nano-sized features has grown because their dimensions are in the same range as the biological components of the bio-recognition layer. A selection of nanostructured biosensors fabricated using ALD is presented in Table 1.

Table 1. Examples of films deposited by ALD for the fabrication of nanostructured biosensors

Material deposited by ALD	Application of the ALD coating	Substrate/Probe	Analyte	Detection Method	Concentration range/Detection Limit	Ref.
Al ₂ O ₃	Sacrificial layer for nanogap fabrication	Gold nanogap/Biotin	Streptavidin	Electrochemical (current)	Detection down to 1.5 nM	(Jan g et al., 2007)
Al ₂ O ₃	Layer for nanoring fabrication	Plasmonic nanoring cavities	Adenine	Optical (surface enhanced Raman spectroscopy [SERS])	Limit of detection 76 nM	(Im et al., 2013)
Al ₂ O ₃	Layer for coax fabrication	Nanoscale coaxial electrodes/anti-cholera toxin antibody	Cholera toxin	Electrochemical (DPV [differential pulse voltammetry] or SWV [square wave voltammetry])	Linear dynamic range of detection of 10 ng/ml – 1 µg/ml Limit of detection 2 ng/ml	(Archi bald et al., 2015)
Al ₂ O ₃	Nano-spacer for metal-enhanced fluorescence (MEF)	Microarray platform on Ag film/split DNA aptamer	17-β-estradiol	Optical (fluorescence)	Detection limit of 1 pg/ml	(Le e et al., 2017)
Al ₂ O ₃ /ZnO	Nanopore size modification, enhancement of sensing capabilities	poly(ethylene terephthalate) (PET) nanochannel/Al ₂ O ₃ /ZnO/biotin	Avidin/streptavidin/bovine serum albumin (BSA)/IgG (Immunoglobulin G)/anti-BSA	Electrochemical (current)	Avidin and streptavidin discrimination/BSA, IgG and anti-BSA detection	(Le poi tevi n et al., 2016)
ZnO	Decoration of silicon nanowires (SiNWs) with ZnO nanoparticles	SiNWs FETs (field effect transistor)/chemical receptors	Explosive chemical species	Electrochemical (FET)	Detection down to the parts-per-quadrillion range	(Lic hte nst ein et al., 2014)
ZnO	Electrode body	ZnO/ZnS core/shell nanotubes/GO _x (glucose oxidase)	Glucose	Electrochemical (current)	Linear range of 2.39 × 10 ⁻⁵ – 2.66 × 10 ⁻⁴ mM	(Ta rish et al., 2017)

Jang *et al.* deposited a film of Al_2O_3 by ALD to be used as a sacrificial layer for the fabrication of a vertical nanogap in a label-free electrical biosensor for the detection of streptavidin (Jang *et al.*, 2007). Al_2O_3 films of 5, 10 and 15 nm were deposited to explore the effect on sensitivity from different gap sizes. Figure 4 shows the vertical nanogap fabricated with 5 nm of Al_2O_3 and 2 nm of Ti. By comparing the ratio of the current before and after binding of streptavidin to biotin, they found out that the 7 nm gap is too small for the binding to take place (no change in current). However, they observed significant current increases for the 12 and 17 nm gaps.




Figure 4. (a) Cross-sectional transmission electron microscopy (TEM) images of a 7 nm vertical gold nanogap. Vertical line is an interfacial boundary of Al_2O_3 and air, which is filled with Pt for TEM analysis and (b) close-up view of circle in (a). Reprinted with permission from (Jang *et al.*, 2007). Copyright 2007 American Vacuum Society.

Similarly, Im *et al.* combined nanosphere lithography (NSL) and ALD for the fabrication of a periodic array of ring-shaped nanocavities with 10 nm gap size to be used as a SERS substrate (Im *et al.*, 2013). Figure 5 presents a schematic representation of the fabrication process. The number of ALD cycles controls the thickness of the Al_2O_3 layer and this, in turn, controls the resulting gap size. Adenine was

used as an analyte to test the film-over-nanospheres (FON) substrate “FON-gap.” SERS measurements revealed a strong 731 cm^{-1} purine stretch coming from adenine that was used to test the detection limits of the SERS substrate. A 76 nM detection limit was obtained, which makes the FON-gap an excellent bioanalytical platform for SERS biosensing.




Figure 5. (a) A schematic representation of the fabrication process for plasmonic nanoring cavities based biosensors, using NSL and ALD. First ALD is used to conformally deposit an Al_2O_3 layer on the FON substrate prepared by depositing Ag films on polystyrene nanospheres dispersed on a glass slide. Another Ag layer is then deposited on the Al_2O_3 layer, forming Ag/ Al_2O_3 /Ag layers stacked on the nanospheres. Anisotropic etching of the top Ag layer using ion-milling reveals the underlying Al_2O_3 layer with a bowl-shaped Ag residual layer on the nanospheres. Partially removing the Al_2O_3 layer reveals Ag/air/Ag nanoring cavities with a nanogap size defined by the thickness of the ALD-grown Al_2O_3 film. (b) Cross-sectional schematics of the fabrication process. (c) Scanning electron micrograph (SEM) of the nanoring cavities on the FON substrates. Scale bar: 200 nm. (d) SEM image of the nanoring cavity array formed over a $16\text{ }\mu\text{m} \times 10\text{ }\mu\text{m}$ area. (e) Photograph of the nanoring cavity (FON-gap) sample. On the standard glass slide, the FON-gap structures are made in a 2 cm-wide circular area. Reprinted with permission from (Im et al., 2013). Copyright 2013 John Wiley and Sons.

In order to fabricate materials with low dimensions, ALD has been used to fill high aspect ratio nanoholes of porous materials (Elam et al., 2006) such as anodic aluminum oxide (AAO) (Banerjee et al., 2009; Yao et al., 2015). Thanks to its controllable pore diameter, periodicity, and density distribution (Poinern et al., 2011), AAO has been incorporated successfully as a template for

biosensors fabrication. As an example, Tarish *et al.* obtained highly ordered ZnO/ZnS core/shell nanotube arrays by using ALD of ZnO and rapid thermal deposition with AAO as a template (Tarish *et al.*, 2017).

ALD has also been used as a tool for tuning single solid-state nanopore sensors. Such sensors, inspired by the pioneering work of Kasianowicz *et al.* on α -hemolysin (Kasianowicz *et al.*, 1996)], are nanometer-sized apertures fabricated in thin films that allow label-free detection of single molecules (Dekker, 2007; Lepoitevin *et al.*, 2017; Miles *et al.*, 2013). Thanks to its conformal deposition, ALD has been able to coat nanopores with high aspect ratio (Spende *et al.*, 2015). For example, Cabello-Aguilar *et al.* deposited Al₂O₃/ZnO nanolaminates to reduce the size of a hydrophobic nanopore (Cabello-Aguilar *et al.*, 2013). The diameter of the nanopore was fine-tuned for α -hemolysin insertions. Al₂O₃/ZnO nanolaminates were chosen due to their low surface roughness, which prevents the collar growth of layers that can clog the nanopore (Balme *et al.*, 2015). The ability of ALD to homogeneously coat the internal surface of a nanopore allows an excellent surface functionalization (Lepoitevin *et al.*, 2015). This capability limits the non-controlled adsorption of proteins outside the track-etched nanopore and is useful for biosensing (Lepoitevin *et al.*, 2016). ALD has also been used to reduce electrical noise in SiN (Chen *et al.*, 2004) and PET (Thangaraj *et al.*, 2016) nanopores by coating them with a thin layer of Al₂O₃.

These examples show the relevance of ALD for the fabrication of nanostructured biosensors and its versatility to be used in combination with other fabrication techniques. The advantages of using ALD to deposit thin films can be seen in the nanometer scale features achieved in the different types of biosensors and the various morphologies that can be coated.

2.2 ALD for the fabrication of transducers

In order to produce a signal that can be measured and correlated to the presence and concentration (in some cases) of a specific analyte, the transducer element of a biosensor requires well-defined morphologies with optimal optical, electrical, chemical, mechanical, and structural properties (Tereshchenko *et al.*, 2016; Velasco-Garcia and Mottram, 2003; Zhang *et al.*, 2000). It is crucial that the interaction between the analyte and the biorecognition element on the transducer material produces a change in one or more of its physicochemical properties (e.g., mass change, photoluminescence change, pH change, photocurrent change). Also, the biocompatibility of the transducer element is necessary to facilitate the immobilization of the biorecognition elements (Guo *et al.*, 2010; Im *et al.*, 2012; Schindler *et al.*, 2008). The capabilities offered by ALD can effectively meet these challenges and examples of oxide transducers fabricated with ALD are presented in Table 2.

Table 2. Examples of materials deposited by ALD for the fabrication of transducers

Material deposited by ALD	Application of the coating	Substrate/Probe	Analyte	Detection method	Concentration range/Detection Limit	Ref.
TiO _x	Transducer layer	Long-period grating (LPG)/TiO _x /Biotin	Avidin	Optical (refractive index)	Detection of avidin-biotin complex	(Dominik <i>et al.</i> , 2017)
ZnO	Transducer layer	ZnO nanorods/GO _x	Glucose	Electrochemical (current)	Sensitivity of 69.8 (nA/ μ M-	(Kim <i>et al.</i> , 2014)

						cm ²)
ZnO	Transducer layer	ZnO film/GVA antibody	GVA antigen	Optical (photoluminescence)	Sensitivity in the range from 1 pg/ml to 10 ng/ml	(Tereshchenko et al., 2017)

Recently, Tereshchenko *et al.* used a ZnO film deposited by ALD as an optical biosensor platform to detect *Grapevine virus A-type* (GVA) proteins (GVA-antigens) for the first time (Tereshchenko et al., 2017). ZnO was chosen due to its high isoelectric point (~9.5), biocompatibility for the immobilization of GVA-antibodies, and ability to change its photoluminescence (PL) emission when interacting with biomolecules. In their work, they deposited a 110 nm thick ZnO layer on a Si substrate at a low deposition temperature of 100 °C. By looking at the PL signal of the biosensor while adding different concentrations of GVA-antigen, they detected the protein with sensitivity in the range from 1 pg/ml to 10 ng/ml. Kim *et al.* grew ZnO nanorods using a hydrothermal method and a seed layer deposited by ALD (Kim et al., 2014). A 20 nm ZnO was deposited and used to promote the growth of high aspect ratio ZnO nanorods in a layer-by-layer fashion. GO_x was immobilized on the ZnO nanorods to detect glucose over a range of different concentrations by cyclic voltammetry. The biosensor showed a sensitivity of 69.8 nA/(μ M·cm²) for the ZnO nanorods with the highest surface area.

Transducer materials can also be functionalized chemically to immobilize the biorecognition element (Arya et al., 2012; Brétangol et al., 2006). The chemical grafting of biomolecules shows an increase of the surface coverage when compared to physical adsorption and improves the general stability and performance of the biosensor (Gervais et al., 2007). TiO_x was deposited on a LPG induced in an optical fiber as a transducer for the recognition of biotin-avidin interactions (Dominik et al., 2017). The transducer layer was functionalized with amine groups to form a peptide bond with the carboxyl group of biotin. X-ray photoelectron spectroscopy (XPS) was used to confirm the successful functionalization with APTES as well as biotin on the TiO_x layer. A 13.2 nm shift in the resonance wavelength was obtained when binding of avidin to biotin took place, proving the biosensor's ability to detect this event. The possibility of fabricating the transducer element in biosensors with ALD has been demonstrated in the examples mentioned above. With the growing need to fabricate biosensors with smaller dimensions, faster response rates, and higher sensitivities, the use of transducers prepared by ALD will become more common and will play a key role in the development of these devices.

2.3 ALD layers to protect biosensors from their environment

ALD deposited films have been used as protective coatings, for example, to limit corrosion processes (Díaz et al., 2013; Shan et al., 2008; Standridge et al., 2009). The ability to coat large areas with substantial uniformity and excellent conformality are some of the advantages of ALD over other thin film deposition techniques. Also, ALD provides dense, pinhole-free films with outstanding adhesion properties (Matero et al., 1999). These benefits have made ALD an excellent choice to protect the surface of biosensors from aqueous environments and render them stable and, in some cases, reusable. To illustrate this, a list of biosensors with protective coatings deposited by ALD is presented in Table 3.

Table 3. Examples of films deposited by ALD for biosensor protection purposes

Material	Application of	Substrate/Probe	Analyte	Detection	Concentration	Ref.
----------	----------------	-----------------	---------	-----------	---------------	------

deposited by ALD	the coating	ALD		Method	range/Detection Limit	
Al ₂ O ₃	Protection against oxidation of Ag	Silver film-over-nanosphere (AgFON)	Anthrax biomarker CaDPA (calcium dipicolinate)	Optical (SERS)	Limit of detection 1.4 x10 ³ spores	(Zhang et al., 2006)
Al ₂ O ₃	Protection of ZnO	ZnO passivated with Al ₂ O ₃	Biofilm	Mass-sensitive (surface acoustic wave [SAW])	Limit of detection 5.3 pg	(Kim et al., 2012)
SiO ₂	Protection of nanolaser structure	InGaAsP multiple quantum well (MQW)/biotin	Streptavidin tagged with phosphor (SAPE)	Optical (air-bridge-type Γ -band-edge laser (BEL) lasing wavelength shift)	Detection of streptavidin-biotin reaction	(Chen et al., 2015)

Cha *et al.* used ALD of SiO₂ to protect the surface of a photonic crystal (PC) BEL structure (Cha et al., 2015). These In-P based structures are subject to chemical attack, and their surface is full of defects that can cause rapid carrier annihilation. By depositing a 5 nm SiO₂ film, they successfully protected the nanolaser device from harsh chemicals and used it as a layer for biotin functionalization. Figure 6 shows the overall functionalization process schematically. Regions with PC pattern show a stronger fluorescence thanks to their higher surface to volume ratio when compared to planar regions. Also, they detected streptavidin with a figure of merit of ~800, which showed the high sensibility of the BEL biosensor.




Figure 6. (a) Schematic representation of the surface functionalization steps for biosensing test. From top-left: 2D PC BEL fabrication; conformal deposition of ALD-SiO₂ layer; biotinylation of silica-terminated surface; and chemical interaction between biotin and streptavidin molecules tagged with

SAPE. (b) Fluorescence microscopy image taken after the dye-conjugated streptavidin was bound onto the biotin-functionalized PC surface. Reprinted with permission from (Cha et al., 2015). Copyright 2015 Royal Society of Chemistry.

The sensitivity of a biosensor can be lowered when a protective layer is applied. The properties of the material to be deposited, as well as its thickness, must be optimized to reduce the possible detrimental effects on the performance of the biosensor. Kim *et al.* used an ALD Al_2O_3 film as a passivation layer to protect the ZnO piezoelectric layer from bacterial growth media or animal serum of a SAW biofilm sensor (Kim et al., 2012). They calculated the normalized theoretical sensitivity of the SAW sensor after applying a protective layer of different materials (Al_2O_3 , Si_3N_4 , SiO_2 , and Teflon) and observed that Al_2O_3 was the one that provided the lowest degradation in sensitivity. A 45 nm Al_2O_3 film was deposited by e-beam evaporation, RF sputtering and ALD to assess the performance of each of the deposition techniques. After two days in Lysogeny Broth (LB) media bacterial suspension, the films prepared by e-beam evaporation and RF sputtering led to some damage and did not protect the ZnO layer successfully as compared to the ALD film (Figure 7). *E. coli* was cultured to test the SAW sensor for biofilm growth monitoring, and a detection limit of 5.3 pg was achieved.




Figure 7. Optical images showing the surface of the SAW sensor passivated with a 45 nm Al_2O_3 film using (a) e-beam evaporation, (b) radio frequency (RF) sputtering (dark area is ZnO) and (c) ALD (no ZnO damage) in LB media with the bacterial solution after two days. Reprinted with permission from (Kim et al., 2012). Copyright 2012 Elsevier.

Zhang *et al.* deposited an ultrathin Al_2O_3 layer on a AgFON substrate for SERS detection (Zhang et al., 2006). The Al_2O_3 layer was used to protect Ag against oxidation and allows preserving an excellent sensitivity. Also, the adsorption affinity of the anthrax biomarker detected in this study for the Al_2O_3 is five times stronger than that for the AgFON alone. The SERS substrate was later used for bacillus spores detection, and a limit of detection (LOD) of $\sim 1.4 \times 10^3$ spores was attained. The SERS substrate remained functional over a period of 9 months and opens up the possibility to be used for biomedical, homeland security, and environmental applications.

These examples illustrate the feasibility of ALD to deposit highly dense and conformal films for the protection of biosensors. The improved long-term stability of ALD coated devices allows them to be reused without having a detrimental impact on their sensing performance and opens up the possibility for biosensors to be used for *in vivo* monitoring.

3. Materials deposited by ALD

The properties of materials at the nanoscale strongly depend on their size, shape, chemical composition, and surface area (Bechelany et al., 2015; S. Zhang et al., 2017). Thus, ALD has been extensively used to modify and enhance different electrical, optical, catalytical, and structural properties of biosensor materials (Chen et al., 2004; Cheun et al., 2010; O'Neill et al., 2015; Wang et al., 2012). In order to be used in biosensing applications, the materials need to be biocompatible with the biological components of the biosensor and shouldn't compromise the sensitivity and overall performance of these devices.

Due to the popularity of FET type biosensors (Li et al., 2018; Sarkar et al., 2014; Zafar et al., 2018), most of the applications of ALD in biosensing have been focused so far on depositing high dielectric materials. Nonetheless, materials with high refractive index and high catalytic activity have also been reported. In this section, the materials deposited by ALD to fabricate biosensors have been classified by their relevant property and are showcased with a few examples.

3.1 Materials with high dielectric constant

ALD has been used to deposit thin films of materials with high dielectric constant to reduce leakage current, passivate surfaces and, in some cases, serve as gate dielectrics (Kim et al., 2009; Ponraj et al., 2013). Examples of high *k* oxides films deposited by ALD for the fabrication of biosensors are presented in Table 4.

Table 4. Examples of materials with high dielectric constant deposited by ALD for the fabrication of biosensors

Material deposited by ALD	Application of the ALD coating	Substrate/Probe	Analyte	Detection method	Concentration range/Detection Limit	Ref.
Al ₂ O ₃	Current leakage prevention	Silicon wire FET/urease	Urea	Electrochemical (FET)	0.1 – 0.68 mM	(Chen et al., 2008)
Al ₂ O ₃ /TiO ₂	Buffer layer during photolithographic process and current leakage decrease/film for ZnO stability	Three-dimensional (3D) bioelectronics-FET (bio-FET)/prostate specific antigen – 1-antichymotrypsin (PSA-ACT) antibody	PSA-ACT antigen	Electrochemical (FET)	Dynamic range of 10 ⁷ and detection down to the fM level	(Kim et al., 2015)
Al ₂ O ₃	Current	Magnetic	Influenza	Magnetolectric	Detection range	(Kri

	leakage prevention	nanoparticles (MNPs)/Influenza A capture antibody	A viruses (IAVs)	(giant magnetoresistance [GMR])	of 1.5×10^2 50% tissue culture infective dose (TCID ₅₀)/ml up to 1.0×10^5 TCID ₅₀ /ml	shna et al., 2016)
Al ₂ O ₃	Insulating layer, current leakage prevention	Nickel interdigitated electrodes/ Al ₂ O ₃ /GSH	glutathione S-transferase (GST)	Electrochemical (impedance)	Lowest detectable value (LDV) 2×10^{-10} M Highest detectable value (HDV) 2×10^{-6} M	(Vello et al., 2017)
Al ₂ O ₃	Insulating and H ⁺ ions sensitive layer	SnO ₂ NW-FET/streptavidin	Biotinylated tetracycline repressor protein (bTetR)	Electrochemical (FET)	Successful detection of 600 µg/ml	(Jakob et al., 2017)
Al ₂ O ₃	Insulating layer, gate dielectric, adhesion improvement	MoS ₂ /anti-PSA antibody	PSA	Electrochemical (FET)	Down to 1 pg/ml	(Yoo et al., 2017)
HfO ₂	Gate dielectric material	Bio-FET	Biotin	Electrochemical (FET)	Biotin functionalization	(Chen et al., 2010)
HfO ₂	Passivation film and gate insulator	Carbon nanotube (CNT) FET/3-Mercaptopropionic acid	Cytochrome C and beta lactoglobulin A	Electrochemical (FET)	Detection range for cytochrome C : 350 pM – 3 µM Detection range for beta lactoglobulin A : 2.5 pM – 21 nM	(Nakashima et al., 2010)
HfO ₂	Reduction of current leakage and enhancer of gate capacitance	Silicon wafer/HfO ₂ /anti-human interleukin-10 (IL-10) monoclonal antibody (mAb)	Recombinant human (rH) IL-10	Electrochemical (impedance)	Linear range of detection of 0.1 – 20 pg/ml	(Lee et al., 2012)
HfO ₂	Gate dielectric layer, protection of electrodes from liquid environment	MoS ₂ nanosheet FET/monoclonal PSA	Cancer marker protein PSA	Electrochemical (FET)	Detection down to 375 fM	(Wang et al., 2014)
HfO ₂	Reduction of leakage current	MoS ₂ FET/anti-human TNF-α	Human TNF-α	Electrochemical (FET)	Detection down to the fM level	(Nam et al., 2011)

ZrO ₂	Gate insulator	Potassium-doped polypyrrole/carbon nanotube/cholesterol oxidase (ChO _x)	Cholesterol	Electrochemical (FET)	Linear range of 0.5 – 20 mM Sensitivity of 400 (μA/mM mm ²)	5) (Barik et al., 2014)
ZrO ₂	Insulating layer	Polydimethylsiloxane (PDMS) microwells/gold electrodes	Normal human dermal fibroblast (NHDF) and human lung adenocarcinoma epithelial (H441)	Electrochemical (impedance)	Successful cell analysis for nanotoxicological studies	(Sticker et al., 2015)

One of the most used high k oxides is Al₂O₃ (Knez et al., 2007; Yan Zhao et al., 2013). Due to its high breakdown field and high thermal stability, Al₂O₃ deposited by ALD can successfully reduce surface leakage current (Ye et al., 2003). Another positive aspect of this material for biosensing is the fact that Al₂O₃ films by ALD present good biocompatibility (Finch et al., 2008). Chen *et al.* deposited a 10 nm film of Al₂O₃ to prevent current leakage between silicon nanowires and analyte solution in a FET biosensor for urea detection (Chen et al., 2008). The Al₂O₃ film was later treated with oxygen plasma (to clean it and render its surface hydrophilic) before functionalization with 3-aminopropyltriethoxy silane (APTES) and urease enzyme. Krishna *et al.* developed a GMR biosensor based on magnetic nanoparticles for detection of influenza A virus (Krishna et al., 2016). In their work, they deposited an 18 nm thick Al₂O₃ film onto a GMR chip to prevent current leakage. Vello *et al.* deposited a 3.3 nm Al₂O₃ insulating film on nickel interdigitated electrodes to prevent leakage current and enable them to operate in an aqueous buffered medium (Vello et al., 2017). To obtain this, they immersed the interdigitated electrodes coated with Al₂O₃ in a phosphate buffered saline (PBS) solution while applying 1 V to the pads of the device. The Al₂O₃ coated electrodes were functionalized with tripeptide reduced glutathione (GSH) to detect the target enzyme GST by evaluating variations on the overall capacitance values. The biosensor could detect GST at concentrations as low as 200 pmol/L (the lowest value reported according to references) and could be easily regenerated to be used several times. Yoo *et al.* deposited an Al₂O₃ layer on an SU-8 organic layer to fabricate a hybrid gate dielectric in a MoS₂ FET biosensor (Yoo et al., 2017). The 30 nm Al₂O₃ layer provided high insulating properties and improved the adhesion between the organic SU-8 layer and the MoS₂/source-drain electrodes multilayer. Figure 8 shows the different layers of the epidermal skin-type MoS₂ biosensor. The flexible biochip was used for real-time detection of PSA with concentrations as low as 1 pg/ml, which is much lower than the value needed for clinical trials (~4 ng/ml). In addition, they introduced a commercial light-emitting diode (LED) to provide a direct diagnostic result, which opens up the possibility to use this type of biosensors in point-of-care (POC) and forensic applications.




Figure 8. Illustration and demonstration of the platform of an epidermal skin-type MoS₂. (a) Schematic of the PSA binding with the PSA antibody functionalized on the MoS₂ surface. (b) and (c) schematic layout of the 2D multilayer MoS₂ FETs. Photograph (d) of an epidermal skin-type MoS₂ biosensor system consisting of the biosensor, read-out circuits, and an LED as an indicator. (e) Optical image and (f) AFM image of a flexible MoS₂ device. Reprinted with permission from (Yoo et al., 2017). Copyright 2017 Springer Nature.

Another high k material that has been deposited by ALD for biosensing applications is HfO₂. HfO₂ is known for its thermodynamic stability when deposited on Si, which effectively reduces leakage current and makes it suitable as a gate material for metal-oxide-semiconductor (MOS) devices (Aarik et al., 2004; Green et al., 2002; Gusev et al., 2003; Gutowski et al., 2002). Also, HfO₂ has an isoelectric point around pH 7 that makes it uncharged in many biological solutions and can be functionalized with biomolecules (Chen et al., 2010). Lee *et al.* developed a biosensor based on HfO₂ for human IL-10 detection by electrochemical impedance spectroscopy (EIS) (Lee et al., 2012). They deposited a 10.7 nm film of HfO₂ on a silicon wafer which was later functionalized with self-assembled monolayers (SAMs) of an aldehyde-silane (11-(triethoxysilyl) undecanal (TESUD)) and anti-human IL-10 mAb. By following the EIS of the modified HfO₂ with increasing human-IL10 concentrations, they demonstrated that the biosensor had a working linear range of 0.1–20 pg/ml and a sensitivity of 0.49 (ng/ml)⁻¹. Wang *et al.* reported a label-free MoS₂ nanosheet-based FET biosensor covered with a 7–8 nm layer of HfO₂ (Wang et al., 2014). Figure 9 shows the homebuilt microfluidic channel system where sample solutions were flowed to be detected by the active area. Figure 6c shows the different layers that make up the FET biosensor. They used HfO₂ to serve as the gate dielectric layer, to protect the metal electrodes from a fluidic environment, and to serve as a starting layer for functionalization with silanes. The FET biosensor showed good sensitivity (down to the femtomolar level) to cancer marker protein PSA and high selectivity by not responding to BSA protein.




Figure 9. Schematic (a) and photo (b) of the biosensor configuration. (c) Schematic of the biofunctionalization layers on the device surface (S [source]; D [drain]). Reprinted with permission from (Wang et al., 2014). Copyright 2014 John Wiley and Sons.

Besides Al_2O_3 and HfO_2 , ALD of ZrO_2 has also been reported for biosensors fabrication. Thanks to its large energy bandgap and thermodynamic stability, ZrO_2 has been considered as an alternative gate dielectric material (Kukli et al., 2002; Nam and Rhee, 2004; Niinistö et al., 2004). Barik *et al.* deposited ZrO_2 by ALD as gate insulator on an enzyme FET (ENFET) biosensor (Barik et al., 2014). They deposited ZrO_2 on the channel region of the biosensor to increase its capacitance and thus gain sensitivity. Sticker *et al.* deposited a 15 nm thick ZrO_2 insulation layer on interdigitated microelectrodes for nanotoxicological cell analysis (Sticker et al., 2015). By running computational simulations, they showed that insulated interdigitated electrode structures had an improved electrical current distribution when compared to bare electrodes. They successfully evaluated the toxicity of silica nanoparticles (with and without protein coatings) on H441 cells by monitoring the impedance signal over time.

The numerous examples of biosensors that employ high k oxides films prepared by ALD highlight the importance of this technique for the fabrication of these devices. Moreover, the conformal and pinhole free thin films of metal oxides successfully passivate the surface of the biosensors without loss of sensitivity and make them promising for miniaturized POC diagnostics.

3.2 Materials with high refractive index

Materials deposited by ALD with high refractive index have been used in optical applications (Alasaarela et al., 2010; Zhu et al., 2016) and microelectromechanical systems (Rissanen et al., 2012). Owing to the smoothness of the films and the ability to coat complex-shaped substrates with a precise thickness control, ALD has been preferred over other thin film deposition techniques like dip coating and self-assembly (Sobel and Hess, 2015; Ying Zhao et al., 2013). Examples of high refractive index films deposited by ALD for the fabrication of biosensors are presented in Table 5. Recently, Oubaha *et al.* deposited a Ta_2O_5 layer by ALD to increase the sensing properties of a multianalyte biosensor (Oubaha et al., 2015). Ta_2O_5 was chosen due to its high refractive index, which increases the intensity of the evanescent field (EF) by 440 times when compared to the system without the high refractive

index layer (HRIL). The thickness of the HRIL chosen for this work was 35 nm because it provides the maximum evanescent wave enhancement. Cy-5-labeled mouse IgG antibody was detected through fluorescence detection, and a limit of detection of 0.25 $\mu\text{g/ml}$ was achieved. Smietana *et al.* deposited an ALD layer of TiO_2 on LPGs for a label-free optical biosensor (Smietana *et al.*, 2015). TiO_2 was used for improving the refractive index sensitivity of the biosensor and as a biocompatible material for endotoxin binding protein (adhesin) functionalization. The deposition of a 70 nm thick TiO_2 layer on the LPG increased the refractive index sensitivity 2.8 times when compared to the bare LPG. Binding of *E. coli* B lipopolysaccharide (LPS) to bacteria adhesion was confirmed thanks to an increase in the spectral separation of resonances.

Table 5. Examples of materials with high refractive index deposited by ALD for the fabrication of biosensors

Material deposited by ALD	Application of the ALD coating	Substrate/Probe	Analyte	Detection Method	Concentration range/Detection Limit	Ref.
Ta_2O_5	EF enhancement	Waveguide-based photonic platform/Cy5-labelled anti-mouse IgG antibody	Mouse IgG antigen	Optical (fluorescence)	Limit of detection of 0.25 $\mu\text{g/ml}$	(Oubaha <i>et al.</i> , 2015)
TiO_2	Improvement of refractive index sensitivity	LPG/ TiO_2 / <i>E. coli</i> B bacteriophage g37 adhesin	<i>E. coli</i> B LPS	Optical (refractive index)	Positive test for adhesin-LPS binding	(Smietana <i>et al.</i> , 2015)

3.3 Materials with catalytic activity

ALD has been used to improve the catalytic activity of materials thanks to its ability to precisely control the catalyst particles properties and uniform dispersion on the surface of the support (Sun *et al.*, 2013; Weber *et al.*, 2015, 2012). Examples of materials with catalytic activity deposited by ALD for the fabrication of biosensors are presented in Table 6. Recently, Choi *et al.* reported the decoration of CNTs with Ni nanoparticles by ALD for non-enzymatic glucose sensing (Choi *et al.*, 2015). The presence of Ni nanoparticles on CNTs was confirmed by TEM, high-angle annular dark-field (HAADF), and energy-dispersive X-ray spectrometry (EDX). The ~ 8 nm Ni nanoparticles were uniformly distributed on the walls of the CNTs. The selectivity of the sensor was confirmed by showing minimum changes in the oxidation current when adding ascorbic acid or uric acid, whereas, in the case of glucose, the value of the oxidation current was far more important. The sensor presented a detection limit of 2 μM and a linear range of 0.005 – 2 mM. Furthermore, it showed a rapid response and repeatability for non-enzymatic detection of glucose.

At the nanoscale, both mass transport and electron transport play a crucial role in obtaining excellent electrocatalytic properties. Nanoporous metals decorated with transition-metal oxides have been proposed as structures that can serve as electrochemical biosensors. Zhang *et al.* deposited a CoO layer on nanoporous gold (NPG) for the detection of glucose and H_2O_2 (Zhang *et al.*, 2016). The Au nanopores with a size ranging between 40-100 nm were covered with 100-800 cycles of CoO. When less than 200 cycles were used, the amorphous CoO formed a discontinuous layer. On the other hand,

a continuous layer was observed at higher cycles. The NPG/CoO heterostructure showed an excellent electrocatalytic activity of glucose oxidation owing to the interconnected Au skeletons and the synergistic effect between Au and CoO. Furthermore, the electrochemical biosensor could detect concentrations of H₂O₂ as low as 0.1 mM and its sensitivity was comparable to other types of composite electrodes based on graphene sheets, noble metals, and metal oxide nanoparticles.

Table 6. Examples of materials with catalytic activity deposited by ALD for the fabrication of biosensors

Material deposited by ALD	Application of the ALD coating	Substrate/Probe	Analyte	Detection Method	Concentration range/Detection Limit	Ref.
CoO	Electrode body, catalytic activity	NPG/CoO	H ₂ O ₂	Electrochemical (current)	Linear range of detection 0.1 – 92.9 mM	(Zhang et al., 2016)
Ni	Electrocatalyst for glucose oxidation	CNT-Ni nanocomposites	Glucose	Electrochemical (current)	Linear range of detection of 0.005 – 2 mM Detection limit of 2 μM	(Choi et al., 2015)
NiO	Electrocatalyst for glucose oxidation	NiO/SiC nanocomposite	Glucose	Electrochemical (current)	Linear range of detection of 0.004 – 7.5 mM Detection limit of 0.32 μM	(Yang et al., 2015)

4. Conclusions and future perspectives

ALD has rapidly become a valuable technique for the fabrication of biosensors thanks to its ability to deposit a wide range of materials with precise thickness control and excellent conformality. In this minireview, the crucial role of ALD for the fabrication of different types of nanostructured biosensors and its capability to improve their sensing properties has been shown. Immobilization of bio-recognition elements has also been demonstrated on layers deposited by ALD thanks to their biocompatibility and controlled chemical composition. The continuous miniaturization trend of biosensor devices and the complexity of the structures to coat demand high conformality, uniformity, and film quality that conventional thin film deposition techniques such as chemical vapor deposition (CVD) and physical vapor deposition (PVD) cannot achieve (Crowell, 2003). However, there are still some challenges to overcome. ALD remains a slow technique and requires vacuum conditions that make the processes somewhat expensive and difficult to scale up (Muñoz-Rojas and MacManus-Driscoll, 2014). Different approaches such as spatial atmospheric ALD (Hoffmann et al., 2018; Hoyer et al., 2015; Poodt et al., 2012), roll-to-roll systems (Ali et al., 2014; Maydannik et al., 2014), and rotating reactors (Longrie et al., 2014; Sharma et al., 2015) are being developed to make the process (even) more scalable and compatible with industrial requirements. Additionally, the design and synthesis of new and affordable precursors are needed to further increase the availability of materials that can be deposited. This is a difficult task given the constraints that the ALD reaction imposes on the precursors. The chemicals must be volatile at room temperature (or by being slightly heated) and

thermally stable. Also, they must react quickly with the substrate to allow surface saturation and provide a fast growth (Johnson et al., 2014; Leskelä and Ritala, 2002; Marichy et al., 2012).

Also, while some biosensors have passed the testing phase and have become available in healthcare applications as handheld devices or portable units (Ronkainen et al., 2010), their design must still be improved so they can diversify their applications in fields such as environmental monitoring, security and bioterrorism, and food safety. Finally, it is safe to say that ALD (in combination with other fabrication techniques) will push biosensors past the limitations they currently face by providing robust, sensible, and selective platforms that will become part of our everyday lives.

5. Acknowledgment

Octavio Graniel would like to thank CONACYT for funding. This work was supported by the European project “CanBioSe, Project ID: 778157”.

List of acronyms

Acronym	Full name
3D	three dimensional
AAO	anodic aluminum oxide
AgFON	silver film-over-nanosphere
ALD	atomic layer deposition
ALE	atomic layer epitaxy
APTES	3-aminopropyltriethoxy silane
BEL	band edge laser
Bio-FET	bioelectronics field effect transistor
BSA	bovine serum albumin
bTetR	biotinylated tetracycline repressor protein
CaDPA	calcium dipicolinate
ChO _x	cholesterol oxidase
CNT	carbon nanotube
CVD	chemical vapor deposition
DPV	differential pulse voltammetry
EDX	energy-dispersive X-ray spectrometry
EF	evanescent field
EIS	electrochemical impedance spectroscopy
ENFET	enzyme field effect transistor
FET	field effect transistor
GMR	giant magnetoresistance
GO _x	glucose oxidase
GSH	tripeptide reduced glutathione
GST	glutathione-S-transferase
GVA	<i>Grapevine virus A-type</i>
H441	human lung adenocarcinoma epithelial
HAADF	high-angle annular dark-field
HDV	highest detectable value
HRIL	high refractive index layer
IAVs	Influenza A viruses
IgG	immunoglobulin G
IL-10	interleukin-10

LB	Lysogeny Broth
LDV	lowest detectable value
LED	light-emitting diode
MBS	3-mal- eimidobenzoic acid N-hydroxysuccinimide ester
ML	molecular layering
MNPs	magnetic nanoparticles
MOS	metal-oxide-semiconductor
MQW	multiple quantum well
NHDF	normal human dermal fibroblast
NPG	nanoporous gold
NSL	nanosphere lithography
PBS	phosphate buffered saline
PC	photonic crystal
PDMS	polydimethylsiloxane
PET	poly(ethylene terephthalate)
POC	point-of-care
PSA	prostate specific antigen
PSA-ACT	prostate specific antigen/1-antichymotrypsin
PVD	physical vapor deposition
SAPE	streptavidin tagged with phosphor
SAW	surface acoustic wave
SWV	square wave voltammetry
TCID	tissue culture infective dose
TEM	transmission electron microscopy
TESUD	11-(triethoxysilyl) undecanal
XPS	(X-ray photoelectron spectroscopy)

Bibliography

- Aaltonen, T., Ritala, M., Tung, Y.L., Chi, Y., Arstila, K., Meinander, K., Leskelä, M., 2004. Atomic layer deposition of noble metals: Exploration of the low limit of the deposition temperature. *J. Mater. Res.* 19, 3353–3358. <https://doi.org/10.1557/JMR.2004.0426>
- Aarik, J., Mändar, H., Kirm, M., Pung, L., 2004. Optical characterization of HfO₂ thin films grown by atomic layer deposition. *Thin Solid Films* 466, 41–47. <https://doi.org/10.1016/j.tsf.2004.01.110>
- Alasaarela, T., Saastamoinen, T., Hiltunen, J., Säynätjoki, A., Tervonen, A., Stenberg, P., Kuittinen, M., Honkanen, S., 2010. Atomic layer deposited titanium dioxide and its application in resonant waveguide grating. *Appl. Opt.* 49, 4321–5. <https://doi.org/10.1364/AO.49.004321>
- Ali, K., Choi, K.-H., Jo, J., Lee, Y.W., 2014. High rate roll-to-roll atmospheric atomic layer deposition of Al₂O₃ thin films towards gas diffusion barriers on polymers. *Mater. Lett.* 136, 90–94. <https://doi.org/10.1016/J.MATLET.2014.07.186>
- Archibald, M.M., Rizal, B., Connolly, T., Burns, M.J., Naughton, M.J., Chiles, T.C., 2015. A nanoaxial-based electrochemical sensor for the detection of cholera toxin. *Biosens. Bioelectron.* 74, 406–410. <https://doi.org/10.1016/j.bios.2015.06.069>
- Arya, S.K., Saha, S., Ramirez-Vick, J.E., Gupta, V., Bhansali, S., Singh, S.P., 2012. Recent advances in ZnO nanostructures and thin films for biosensor applications: Review. *Anal. Chim. Acta* 737, 1–21. <https://doi.org/10.1016/j.aca.2012.05.048>
- Balme, S., Picaud, F., Manghi, M., Palmeri, J., Bechelany, M., Cabello-Aguilar, S., Abou-Chaaya, A., Miele, P., Balanzat, E., Janot, J.M., 2015. Ionic transport through sub-10nm diameter hydrophobic

- high-aspect ratio nanopores: Experiment, theory and simulation. *Sci. Rep.* 5, 10135.
<https://doi.org/10.1038/srep10135>
- Banerjee, P., Perez, I., Henn-Lecordier, L., Lee, S.B., Rubloff, G.W., 2009. Nanotubular metal-insulator-metal capacitor arrays for energy storage. *Nat. Nanotechnol.* 4, 292–296.
<https://doi.org/10.1038/nnano.2009.37>
- Barik, M.A., Sarma, M.K., Sarkar, C.R., Dutta, J.C., 2014. Highly Sensitive Potassium-Doped Polypyrrole/Carbon Nanotube-Based Enzyme Field Effect Transistor (ENFET) for Cholesterol Detection. *Appl. Biochem. Biotechnol.* 174, 1104–1114. <https://doi.org/10.1007/s12010-014-1029-5>
- Bechelany, M., Balme, S., Miele, P., 2015. Atomic layer deposition of biobased nanostructured interfaces for energy, environmental and health applications. *Pure Appl. Chem.* 87, 751–758.
<https://doi.org/10.1515/pac-2015-0102>
- Black, A., Urbanos, F.J., Osorio, M.R., Miranda, R., Vázquez De Parga, A.L., Granados, D., 2018. Encapsulating Chemically Doped Graphene via Atomic Layer Deposition. *ACS Appl. Mater. Interfaces* 10, 8190–8196. <https://doi.org/10.1021/acsami.7b18709>
- Brétangol, F., Valsesia, A., Ceccone, G., Colpo, P., Gilliland, D., Ceriotti, L., Hasiwa, M., Rossi, F., 2006. Surface functionalization and patterning techniques to design interfaces for biomedical and biosensor applications. *Plasma Process. Polym.* 3, 443–455.
<https://doi.org/10.1002/ppap.200600015>
- Cabello-Aguilar, S., Balme, S., Chaaya, A.A., Bechelany, M., Balanzat, E., Janot, J.M., Pochat-Bohatier, C., Miele, P., Dejardin, P., 2013. Slow translocation of polynucleotides and their discrimination by α -hemolysin inside a single track-etched nanopore designed by atomic layer deposition. *Nanoscale* 5, 9582–9586. <https://doi.org/10.1039/c3nr03683a>
- Cha, H., Lee, J., Jordan, L.R., Lee, S.H., Oh, S.-H., Kim, H.J., Park, J., Hong, S., Jeon, H., 2015. Surface passivation of a photonic crystal band-edge laser by atomic layer deposition of SiO₂ and its application for biosensing. *Nanoscale* 7, 3565–3571. <https://doi.org/10.1039/C4NR07552H>
- Chaaya, A.A., Viter, R., Baleviciute, I., Bechelany, M., Ramanavicius, A., Gertner, Z., Erts, D., Smyntyna, V., Miele, P., 2014. Tuning Optical Properties of Al₂O₃/ZnO Nanolaminates Synthesized by Atomic Layer Deposition. *J. Phys. Chem. C* 118, 3811–3819.
<https://doi.org/10.1021/jp411970w>
- Chalker, P.R., 2016. Photochemical atomic layer deposition and etching. *Surf. Coatings Technol.* 291, 258–263. <https://doi.org/10.1016/j.surfcoat.2016.02.046>
- Chen, P., Mitsui, T., Farmer, D.B., Golovchenko, J., Gordon, R.G., Branton, D., 2004. Atomic layer deposition to fine-tune the surface properties and diameters of fabricated nanopores. *Nano Lett.* 4, 1333–1337. <https://doi.org/10.1021/nl0494001>
- Chen, Y., Wang, X., Hong, M., Erramilli, S., Mohanty, P., 2008. Surface-modified silicon nano-channel for urea sensing. *Sensors Actuators B Chem.* 133, 593–598.
<https://doi.org/10.1016/j.snb.2008.03.033>
- Chen, Y.W., Liu, M., Kaneko, T., McIntyre, P.C., 2010. Atomic Layer Deposited Hafnium Oxide Gate Dielectrics for Charge-Based Biosensors. *Electrochem. Solid-State Lett.* 13, G29–G32.
<https://doi.org/10.1149/1.3280224>
- Cheun, H., Fuentes-Hernandez, C., Zhou, Y., Potscavage, W.J., Kim, S.J., Shim, J., Dindar, A., Kippelen, B., 2010. Electrical and optical properties of ZnO processed by atomic layer deposition

- in inverted polymer solar cells. *J. Phys. Chem. C* 114, 20713–20718.
<https://doi.org/10.1021/jp106641j>
- Choi, T., Kim, S.H., Lee, C.W., Kim, H., Choi, S.K., Kim, S.H., Kim, E., Park, J., Kim, H., 2015. Synthesis of carbon nanotube-nickel nanocomposites using atomic layer deposition for high-performance non-enzymatic glucose sensing. *Biosens. Bioelectron.* 63, 325–330.
<https://doi.org/10.1016/j.bios.2014.07.059>
- Crowell, J.E., 2003. Chemical methods of thin film deposition: Chemical vapor deposition, atomic layer deposition, and related technologies. *J. Vac. Sci. Technol. A Vacuum, Surfaces, Film.* 21, S88–S95. <https://doi.org/10.1116/1.1600451>
- Dekker, C., 2007. Solid-state nanopores. *Nat. Nanotechnol.* 2, 209–215.
<https://doi.org/10.1038/nnano.2007.27>
- Díaz, B., Światowska, J., Maurice, V., Seyeux, A., Härkönen, E., Ritala, M., Tervakangas, S., Kolehmainen, J., Marcus, P., 2013. Tantalum oxide nanocoatings prepared by atomic layer and filtered cathodic arc deposition for corrosion protection of steel: Comparative surface and electrochemical analysis. *Electrochim. Acta* 90, 232–245.
<https://doi.org/10.1016/j.electacta.2012.12.007>
- Dominik, M., Leśniewski, A., Janczuk, M., Niedziółka-Jönsson, J., Hołdyński, M., Wachnicki, Ł., Godlewski, M., Bock, W.J., Śmietana, M., 2017. Titanium oxide thin films obtained with physical and chemical vapour deposition methods for optical biosensing purposes. *Biosens. Bioelectron.* 93, 102–109. <https://doi.org/10.1016/j.bios.2016.09.079>
- Elam, J.W., Xiong, G., Han, C.Y., Wang, H.H., Birrell, J.P., Welp, U., Hryn, J.N., Pellin, M.J., Baumann, T.F., Poco, J.F., Satcher, J.H., 2006. Atomic layer deposition for the conformal coating of nanoporous materials. *J. Nanomater.* 2006, 1–5. <https://doi.org/10.1155/JNM/2006/64501>
- Elias, J., Bechelany, M., Utke, I., Erni, R., Hosseini, D., Michler, J., Philippe, L., 2012. Urchin-inspired zinc oxide as building blocks for nanostructured solar cells. *Nano Energy* 1, 696–705.
<https://doi.org/10.1016/j.nanoen.2012.07.002>
- Finch, D.S., Oreskovic, T., Ramadurai, K., Herrmann, C.F., George, S.M., Mahajan, R.L., 2008. Biocompatibility of atomic layer-deposited alumina thin films. *J. Biomed. Mater. Res. - Part A* 87, 100–106. <https://doi.org/10.1002/jbm.a.31732>
- Fu, X., Cheng, Z., Yu, J., Choo, P., Chen, L., Choo, J., 2016. A SERS-based lateral flow assay biosensor for highly sensitive detection of HIV-1 DNA. *Biosens. Bioelectron.* 78, 530–537.
<https://doi.org/10.1016/J.BIOS.2015.11.099>
- George, S.M., Steven M. George, 2010. Atomic Layer Deposition: An Overview. *Chem. Rev.* 110, 111–131. <https://doi.org/10.1021/cr900056b>
- Gervais, L., Gel, M., Allain, B., Tolba, M., Brovko, L., Zourob, M., Mandeville, R., Griffiths, M., Evoy, S., 2007. Immobilization of biotinylated bacteriophages on biosensor surfaces. *Sensors Actuators, B Chem.* 125, 615–621. <https://doi.org/10.1016/j.snb.2007.03.007>
- Gong, Y., Palacio, D., Song, X., Patel, R.L., Liang, X., Zhao, X., Goodenough, J.B., Huang, K., 2013. Stabilizing nanostructured solid oxide fuel cell cathode with atomic layer deposition. *Nano Lett.* 13, 4340–4345. <https://doi.org/10.1021/nl402138w>
- Green, M.L., Ho, M.-Y., Busch, B., Wilk, G.D., Sorsch, T., Conard, T., Brijs, B., Vandervorst, W., Räisänen, P.I., Muller, D., Bude, M., Grazul, J., 2002. Nucleation and growth of atomic layer

- deposited HfO₂ gate dielectric layers on chemical oxide (Si–O–H) and thermal oxide (SiO₂ or Si–O–N) underlayers. *J. Appl. Phys.* 92, 7168–7174. <https://doi.org/10.1063/1.1522811>
- Gu, Y.-Z., Lu, H., Geng, Y., Ye, Z.-Y., Zhang, Y., Sun, Q.-Q., Ding, S.-J., Zhang, D.W., 2013. Optical and microstructural properties of ZnO/TiO₂ nanolaminates prepared by atomic layer deposition. *Nanoscale Res. Lett.* 8, 107. <https://doi.org/10.1186/1556-276X-8-107>
- Guo, D.J., Abdulagatov, A.I., Rourke, D.M., Bertness, K.A., George, S.M., Lee, Y.C., Tan, W., 2010. GaN nanowire functionalized with atomic layer deposition techniques for enhanced immobilization of biomolecules. *Langmuir* 26, 18382–18391. <https://doi.org/10.1021/la103337a>
- Gusev, E.P., Copel, M., Gribelyuk, M., 2003. Ultrathin HfO₂ films grown on silicon by atomic layer deposition for advanced gate dielectrics applications. *Microelectron. Eng.* 69, 145–151. [https://doi.org/10.1016/S0167-9317\(03\)00291-0](https://doi.org/10.1016/S0167-9317(03)00291-0)
- Gutowski, M., Jaffe, J.E., Liu, C.-L., Stoker, M., Hegde, R.I., Rai, R.S., Tobin, P.J., 2002. Thermodynamic stability of high-K dielectric metal oxides ZrO₂ and HfO₂ in contact with Si and SiO₂. *Appl. Phys. Lett.* 80, 1897–1899. <https://doi.org/10.1063/1.1458692>
- Hämäläinen, J., Ritala, M., Leskelä, M., 2014. Atomic Layer Deposition of Noble Metals and Their Oxides. *Chem. Mater.* 26, 786–801. <https://doi.org/10.1021/cm402221y>
- Ho, T.A., Bae, C., Nam, H., Kim, E., Lee, S.Y., Park, J.H., Shin, H., 2018. Metallic Ni₃S₂ Films Grown by Atomic Layer Deposition as an Efficient and Stable Electrocatalyst for Overall Water Splitting. *ACS Appl. Mater. Interfaces* 10, 12807–12815. <https://doi.org/10.1021/acsami.8b00813>
- Hoffmann, L., Brinkmann, K.O., Malerczyk, J., Rogalla, D., Becker, T., Theirich, D., Shutsko, I., Görrn, P., Riedl, T., 2018. Spatial Atmospheric Pressure Atomic Layer Deposition of Tin Oxide as an Impermeable Electron Extraction Layer for Perovskite Solar Cells with Enhanced Thermal Stability. *ACS Appl. Mater. Interfaces* 10, 6006–6013. <https://doi.org/10.1021/acsami.7b17701>
- Hong, K.J., Kim, S.O., 2016. Atomic layer deposition assisted sacrificial template synthesis of mesoporous TiO₂ electrode for high performance lithium ion battery anodes. *Energy Storage Mater.* 2, 27–34. <https://doi.org/10.1016/j.ensm.2015.11.002>
- Hong, Y., Kim, C.H., Shin, J., Kim, K.Y., Kim, J.S., Hwang, C.S., Lee, J.H., 2016. Highly selective ZnO gas sensor based on MOSFET having a horizontal floating-gate. *Sensors Actuators, B Chem.* 232, 653–659. <https://doi.org/10.1016/j.snb.2016.04.010>
- Hoye, R.L.Z., Muñoz-Rojas, D., Nelson, S.F., Illiberi, A., Poodt, P., Roozeboom, F., MacManus-Driscoll, J.L., 2015. Research Update: Atmospheric pressure spatial atomic layer deposition of ZnO thin films: Reactors, doping, and devices. *APL Mater.* 3, 40701. <https://doi.org/10.1063/1.4916525>
- Im, H., Bantz, K.C., Lee, S.H., Johnson, T.W., Haynes, C.L., Oh, S.-H., 2013. Self-Assembled Plasmonic Nanoring Cavity Arrays for SERS and LSPR Biosensing. *Adv. Mater.* 25, 2678–2685. <https://doi.org/10.1002/adma.201204283>
- Im, H., Wittenberg, N.J., Lindquist, N.C., Oh, S.H., 2012. Atomic layer deposition: A versatile technique for plasmonics and nanobiotechnology. *J. Mater. Res.* <https://doi.org/10.1557/jmr.2011.434>
- Jakob, M.H., Dong, B., Gutsch, S., Chatelle, C., Krishnaraja, A., Weber, W., Zacharias, M., 2017. Label-free SnO₂ nanowire FET biosensor for protein detection. *Nanotechnology* 28, 245503. <https://doi.org/10.1088/1361-6528/aa7015>

- Jang, D.-Y., Kim, Y.-P., Kim, H.-S., Ko Park, S.-H., Choi, S.-Y., Choi, Y.-K., 2007. Sublithographic vertical gold nanogap for label-free electrical detection of protein-ligand binding. *J. Vac. Sci. Technol. B Microelectron. Nanom. Struct.* 25, 443. <https://doi.org/10.1116/1.2713403>
- Johnson, R.W., Hultqvist, A., Bent, S.F., 2014. A brief review of atomic layer deposition: From fundamentals to applications. *Mater. Today* 17, 236–246. <https://doi.org/10.1016/j.mattod.2014.04.026>
- Kasianowicz, J.J., Brandin, E., Branton, D., Deamer, D.W., 1996. Characterization of individual polynucleotide molecules using a membrane channel. *Proc. Natl. Acad. Sci.* 93, 13770–13773. <https://doi.org/10.1073/pnas.93.24.13770>
- Kim, B.-Y., Sohn, I., Lee, D., Han, G.S., Lee, W.-I., Jung, H.S., Lee, N.-E., 2015. Ultrarapid and ultrasensitive electrical detection of proteins in a three-dimensional biosensor with high capture efficiency. *Nanoscale* 7, 9844–9851. <https://doi.org/10.1039/C5NR00909J>
- Kim, H., 2003. Atomic layer deposition of metal and nitride thin films: Current research efforts and applications for semiconductor device processing. *J. Vac. Sci. Technol. B Microelectron. Nanom. Struct.* 21, 2231. <https://doi.org/10.1116/1.1622676>
- Kim, H., Lee, H.B.R., Maeng, W.J., 2009. Applications of atomic layer deposition to nanofabrication and emerging nanodevices. *Thin Solid Films* 517, 2563–2580. <https://doi.org/10.1016/j.tsf.2008.09.007>
- Kim, J.Y., Jo, S.-Y., Sun, G.-J., Katoch, A., Choi, S.-W., Kim, S.S., 2014. Tailoring the surface area of ZnO nanorods for improved performance in glucose sensors. *Sensors Actuators B Chem.* 192, 216–220. <https://doi.org/10.1016/j.snb.2013.10.113>
- Kim, Y.W., Sardari, S.E., Meyer, M.T., Iliadis, A.A., Wu, H.C., Bentley, W.E., Ghodssi, R., 2012. An ALD aluminum oxide passivated Surface Acoustic Wave sensor for early biofilm detection. *Sensors Actuators, B Chem.* 163, 136–145. <https://doi.org/10.1016/j.snb.2012.01.021>
- Knez, M., Nielsch, K., Niinistö, L., 2007. Synthesis and surface engineering of complex nanostructures by atomic layer deposition. *Adv. Mater.* 19, 3425–3438. <https://doi.org/10.1002/adma.200700079>
- Krishna, V.D., Wu, K., Perez, A.M., Wang, J.-P., 2016. Giant Magnetoresistance-based Biosensor for Detection of Influenza A Virus. *Front. Microbiol.* 7, 400. <https://doi.org/10.3389/fmicb.2016.00400>
- Kukli, K., Ritala, M., Aarik, J., Uustare, T., Leskelä, M., 2002. Influence of growth temperature on properties of zirconium dioxide films grown by atomic layer deposition. *J. Appl. Phys.* 92, 1833–1840. <https://doi.org/10.1063/1.1493657>
- Lee, M., Zine, N., Baraket, A., Zabala, M., Campabadal, F., Caruso, R., Trivella, M.G., Jaffrezic-Renault, N., Errachid, A., 2012. A novel biosensor based on hafnium oxide: Application for early stage detection of human interleukin-10. *Sensors Actuators, B Chem.* 175, 201–207. <https://doi.org/10.1016/j.snb.2012.04.090>
- Lee, W.-I., Shrivastava, S., Duy, L.-T., Yeong Kim, B., Son, Y.-M., Lee, N.-E., 2017. A smartphone imaging-based label-free and dual-wavelength fluorescent biosensor with high sensitivity and accuracy. *Biosens. Bioelectron.* 94, 643–650. <https://doi.org/10.1016/j.bios.2017.03.061>
- Lepoitevin, M., Bechelany, M., Balanzat, E., Janot, J.M., Balme, S., 2016. Non-Fluorescence label protein sensing with track-etched nanopore decorated by avidin/biotin system. *Electrochim. Acta* 211, 611–618. <https://doi.org/10.1016/j.electacta.2016.06.079>

- Lepoitevin, M., Ma, T., Bechelany, M., Janot, J.-M., Balme, S., 2017. Functionalization of single solid state nanopores to mimic biological ion channels: A review. *Adv. Colloid Interface Sci.* 250, 195–213. <https://doi.org/10.1016/j.cis.2017.09.001>
- Lepoitevin, M., Nguyen, G., Bechelany, M., Balanzat, E., Janot, J.M., Balme, S., 2015. Combining a sensor and a pH-gated nanopore based on an avidin-biotin system. *Chem. Commun.* 51, 5994–5997. <https://doi.org/10.1039/c4cc10087e>
- Leskelä, M., Ritala, M., 2003. Atomic Layer Deposition Chemistry: Recent Developments and Future Challenges. *Angew. Chemie Int. Ed.* 42, 5548–5554. <https://doi.org/10.1002/anie.200301652>
- Leskelä, M., Ritala, M., 2002. Atomic layer deposition (ALD): From precursors to thin film structures, in: *Thin Solid Films*. Elsevier, pp. 138–146. [https://doi.org/10.1016/S0040-6090\(02\)00117-7](https://doi.org/10.1016/S0040-6090(02)00117-7)
- Li, Q., Lu, N., Wang, L., Fan, C., 2018. Advances in Nanowire Transistor-Based Biosensors. *Small Methods* 2, 1700263. <https://doi.org/10.1002/smtd.201700263>
- Li, X., Puttaswamy, M., Wang, Z., Kei Tan, C., Grimsdale, A.C., Kherani, N.P., Tok, A.I.Y., 2017. A pressure tuned stop-flow atomic layer deposition process for MoS₂ on high porous nanostructure and fabrication of TiO₂/MoS₂ core/shell inverse opal structure. *Appl. Surf. Sci.* 422, 536–543. <https://doi.org/10.1016/j.apsusc.2017.06.016>
- Lichtenstein, A., Havivi, E., Shacham, R., Hahamy, E., Leibovich, R., Pevzner, A., Krivitsky, V., Davivi, G., Presman, I., Elnathan, R., Engel, Y., Flaxer, E., Patolsky, F., 2014. Supersensitive fingerprinting of explosives by chemically modified nanosensors arrays. *Nat. Commun.* 5, 4195. <https://doi.org/10.1038/ncomms5195>
- Lim, B.S., Rahtu, A., Gordon, R.G., 2003. Atomic layer deposition of transition metals. *Nat. Mater.* 2, 749–754. <https://doi.org/10.1038/nmat1000>
- Liu, J., Banis, M.N., Sun, Q., Lushington, A., Li, R., Sham, T.K., Sun, X., 2014. Rational Design of Atomic-Layer-Deposited LiFePO₄ as a High-Performance Cathode for Lithium-Ion Batteries. *Adv. Mater.* 26, 6472–6477. <https://doi.org/10.1002/adma.201401805>
- Liu, S., Tong, Z., Mu, X., Liu, B., Du, B., Liu, Z., Gao, C., 2018. Detection of Abrin by Electrochemiluminescence Biosensor Based on Screen Printed Electrode. *Sensors* 18, 357. <https://doi.org/10.3390/s18020357>
- Longrie, D., Deduytsche, D., Haemers, J., Smet, P.F., Driesen, K., Detavernier, C., 2014. Thermal and Plasma-Enhanced Atomic Layer Deposition of TiN Using TDMAT and NH₃ on Particles Agitated in a Rotary Reactor. *ACS Appl. Mater. Interfaces* 6, 7316–7324. <https://doi.org/10.1021/am5007222>
- Luo, J., Liu, X., Tian, Q., Yue, W., Zeng, J., Chen, G., Cai, X., 2009. Disposable bioluminescence-based biosensor for detection of bacterial count in food. *Anal. Biochem.* 394, 1–6. <https://doi.org/10.1016/j.ab.2009.05.021>
- Malygin, A.A., Drozd, V.E., Malkov, A.A., Smirnov, V.M., 2015. From V. B. Aleskovskii's "Framework" Hypothesis to the Method of Molecular Layering/Atomic Layer Deposition. *Chem. Vap. Depos.* 21, 216–240. <https://doi.org/10.1002/cvde.201502013>
- Marichy, C., Bechelany, M., Pinna, N., 2012. Atomic layer deposition of nanostructured materials for energy and environmental applications. *Adv. Mater.* 24, 1017–1032. <https://doi.org/10.1002/adma.201104129>

- Matero, R., Ritala, M., Leskelä, M., Salo, T., Aromaa, J., Forsén, O., 1999. Atomic layer deposited thin films for corrosion protection. *Le J. Phys. IV* 9, Pr8-493-Pr8-499. <https://doi.org/10.1051/jp4:1999862>
- Maydannik, P.S., Kääriäinen, T.O., Lahtinen, K., Cameron, D.C., Söderlund, M., Soininen, P., Johansson, P., Kuusipalo, J., Moro, L., Zeng, X., 2014. Roll-to-roll atomic layer deposition process for flexible electronics encapsulation applications. *J. Vac. Sci. Technol. A Vacuum, Surfaces, Film.* 32, 51603. <https://doi.org/10.1116/1.4893428>
- Meng, X., Cao, Y., Libera, J.A., Elam, J.W., 2017. Atomic Layer Deposition of Aluminum Sulfide: Growth Mechanism and Electrochemical Evaluation in Lithium-Ion Batteries. *Chem. Mater.* 29, 9043–9052. <https://doi.org/10.1021/acs.chemmater.7b02175>
- Miles, B.N., Ivanov, A.P., Wilson, K.A., Doğan, F., Japrun, D., Edel, J.B., 2013. Single molecule sensing with solid-state nanopores: Novel materials, methods, and applications. *Chem. Soc. Rev.* <https://doi.org/10.1039/c2cs35286a>
- Mishra, A., Kumar, J., Melo, J.S., 2017. An optical microplate biosensor for the detection of methyl parathion pesticide using a biohybrid of *Sphingomonas* sp. cells-silica nanoparticles. *Biosens. Bioelectron.* 87, 332–338. <https://doi.org/10.1016/j.BIOS.2016.08.048>
- Muhammad-Tahir, Z., Alocilja, E.C., 2003. A conductometric biosensor for biosecurity. *Biosens. Bioelectron.* 18, 813–819. [https://doi.org/10.1016/S0956-5663\(03\)00020-4](https://doi.org/10.1016/S0956-5663(03)00020-4)
- Muñoz-Rojas, D., MacManus-Driscoll, J., 2014. Spatial atmospheric atomic layer deposition: a new laboratory and industrial tool for low-cost photovoltaics. *Mater. Horiz.* 1, 314–320. <https://doi.org/10.1039/C3MH00136A>
- Nakashima, Y., Ohno, Y., Kishimoto, S., Okochi, M., Honda, H., Mizutani, T., 2010. Fabrication Process of Carbon Nanotube Field Effect Transistors Using Atomic Layer Deposition Passivation for Biosensors. *J. Nanosci. Nanotechnol.* 10, 3805–3809. <https://doi.org/10.1166/jnn.2010.1983>
- Nam, H., Oh, B.-R., Chen, P., Yoon, J.S., Wi, S., Chen, M., Kurabayashi, K., Liang, X., 2015. Two different device physics principles for operating MoS₂ transistor biosensors with femtomolar-level detection limits. *Appl. Phys. Lett.* 107, 12105. <https://doi.org/10.1063/1.4926800>
- Nam, W.H., Rhee, S.W., 2004. Atomic layer deposition of ZrO₂ thin films using dichlorobis[bis-(trimethylsilyl)amido]zirconium and water. *Chem. Vap. Depos.* 10, 201–205. <https://doi.org/10.1002/cvde.200306277>
- Niinistö, L., Päiväsaari, J., Niinistö, J., Putkonen, M., Nieminen, M., 2004. Advanced electronic and optoelectronic materials by atomic layer deposition: An overview with special emphasis on recent progress in processing of high-k dielectrics and other oxide materials. *Phys. Status Solidi Appl. Res.* <https://doi.org/10.1002/pssa.200406798>
- O'Neill, B.J., Jackson, D.H.K., Lee, J., Canlas, C., Stair, P.C., Marshall, C.L., Elam, J.W., Kuech, T.F., Dumesic, J.A., Huber, G.W., 2015. Catalyst Design with Atomic Layer Deposition. *ACS Catal.* 5, 1804–1825. <https://doi.org/10.1021/cs501862h>
- Oubaha, M., Gorin, A., McDonagh, C., Duffy, B., Copperwhite, R., 2015. Development of a multianalyte optical sol-gel biosensor for medical diagnostic. *Sensors Actuators B Chem.* 221, 96–103. <https://doi.org/10.1016/j.snb.2015.06.012>
- Pessoa, R.S., dos Santos, V.P., Cardoso, S.B., Doria, A.C.O.C., Figueira, F.R., Rodrigues, B.V.M., Testoni, G.E., Fraga, M.A., Marciano, F.R., Lobo, A.O., Maciel, H.S., 2017. TiO₂ coatings via

- atomic layer deposition on polyurethane and polydimethylsiloxane substrates: Properties and effects on *C. albicans* growth and inactivation process. *Appl. Surf. Sci.* 422, 73–84.
<https://doi.org/10.1016/j.apsusc.2017.05.254>
- Peters, A.W., Li, Z., Farha, O.K., Hupp, J.T., 2015. Atomically Precise Growth of Catalytically Active Cobalt Sulfide on Flat Surfaces and within a Metal–Organic Framework via Atomic Layer Deposition. *ACS Nano* 9, 8484–8490. <https://doi.org/10.1021/acsnano.5b03429>
- Poinern, G.E.J., Ali, N., Fawcett, D., 2011. Progress in Nano-Engineered Anodic Aluminum Oxide Membrane Development. *Materials* 4, 487–526. <https://doi.org/10.3390/ma4030487>
- Ponraj, J.S., Attolini, G., Bosi, M., 2013. Review on Atomic Layer Deposition and Applications of Oxide Thin Films. *Crit. Rev. Solid State Mater. Sci.* 38, 203–233.
<https://doi.org/10.1080/10408436.2012.736886>
- Poodt, P., Cameron, D.C., Dickey, E., George, S.M., Kuznetsov, V., Parsons, G.N., Roozeboom, F., Sundaram, G., Vermeer, A., 2012. Spatial atomic layer deposition: A route towards further industrialization of atomic layer deposition. *J. Vac. Sci. Technol. A Vacuum, Surfaces, Film.* 30, 10802. <https://doi.org/10.1116/1.3670745>
- Puurunen, R.L., 2014. A short history of atomic layer deposition: Tuomo Suntola's atomic layer epitaxy. *Chem. Vap. Depos.* 20, 332–344. <https://doi.org/10.1002/cvde.201402012>
- Rissanen, A., Kantojärvi, U., Blomberg, M., Antila, J., Eränen, S., 2012. Monolithically integrated microspectrometer-on-chip based on tunable visible light MEMS FPI. *Sensors Actuators A Phys.* 182, 130–135. <https://doi.org/10.1016/j.sna.2012.05.023>
- Ritala, M., Leskelä, M., 2002. Atomic layer deposition, in: *Handbook of Thin Films*. Elsevier, pp. 103–159. <https://doi.org/10.1016/B978-012512908-4/50005-9>
- Ronkainen, N.J., Halsall, H.B., Heineman, W.R., 2010. Electrochemical biosensors. *Chem. Soc. Rev.* 39, 1747–1763. <https://doi.org/10.1039/b714449k>
- Sarkar, D., Liu, W., Xie, X., Anselmo, A.C., Mitragotri, S., Banerjee, K., 2014. MoS₂ Field-Effect Transistor for Next-Generation Label-Free Biosensors. *ACS Nano* 8, 3992–4003.
<https://doi.org/10.1021/nn5009148>
- Schindler, M., Kim, S.K., Hwang, C.S., Schindler, C., Offenhäusser, A., Ingebrandt, S., 2008. Novel post-process for the passivation of a CMOS biosensor. *Phys. status solidi – Rapid Res. Lett.* 2, 4–6. <https://doi.org/10.1002/pssr.200701242>
- Shan, C.X., Hou, X., Choy, K.L., 2008. Corrosion resistance of TiO₂ films grown on stainless steel by atomic layer deposition. *Surf. Coatings Technol.* 202, 2399–2402.
<https://doi.org/10.1016/j.surfcoat.2007.08.066>
- Sharma, K., Hall, R.A., George, S.M., 2015. Spatial atomic layer deposition on flexible substrates using a modular rotating cylinder reactor. *J. Vac. Sci. Technol. A Vacuum, Surfaces, Film.* 33, 01A132.
<https://doi.org/10.1116/1.4902086>
- Smietana, M., Koba, M., Brzozowska, E., Krogulski, K., Nakonieczny, J., Wachnicki, L., Mikulic, P., Godlewski, M., Bock, W.J., 2015. Label-free sensitivity of long-period gratings enhanced by atomic layer deposited TiO₂ nano-overlays. *Opt. Express* 23, 8441.
<https://doi.org/10.1364/OE.23.008441>
- Sobel, N., Hess, C., 2015. *Nanoscale Structuring of Surfaces by Using Atomic Layer Deposition*.

Angew. Chemie Int. Ed. 54, 15014–15021. <https://doi.org/10.1002/anie.201503680>

- Spende, A., Sobel, N., Lukas, M., Zierold, R., Riedl, J.C., Gura, L., Schubert, I., Moreno, J.M.M., Nielsch, K., Stühn, B., Hess, C., Trautmann, C., Toimil-Molares, M.E., 2015. TiO₂, SiO₂, and Al₂O₃ coated nanopores and nanotubes produced by ALD in etched ion-track membranes for transport measurements. *Nanotechnology* 26, 335301. <https://doi.org/10.1088/0957-4484/26/33/335301>
- Standridge, S.D., Schatz, G.C., Hupp, J.T., 2009. Toward plasmonic solar cells: Protection of silver nanoparticles via atomic layer deposition of TiO₂. *Langmuir* 25, 2596–2600. <https://doi.org/10.1021/la900113e>
- Sticker, D., Rothbauer, M., Charwat, V., Steinkühler, J., Bethge, O., Bertagnolli, E., Wanzenboeck, H.D., Ertl, P., 2015. Zirconium dioxide nanolayer passivated impedimetric sensors for cell-based assays. *Sensors Actuators, B Chem.* 213, 35–44. <https://doi.org/10.1016/j.snb.2015.02.018>
- Sun, S., Zhang, G., Gauquelin, N., Chen, N., Zhou, J., Yang, S., Chen, W., Meng, X., Geng, D., Banis, M.N., Li, R., Ye, S., Knights, S., Botton, G.A., Sham, T.K., Sun, X., 2013. Single-atom catalysis using Pt/graphene achieved through atomic layer deposition. *Sci. Rep.* 3, 1775. <https://doi.org/10.1038/srep01775>
- Suntola, T., Antson, J., 1977. Method for producing compound thin films. US Pat. 4,058,430.
- Tarish, S., Xu, Y., Wang, Z., Mate, F., Al-Haddad, A., Wang, W., Lei, Y., 2017. Highly efficient biosensors by using well-ordered ZnO/ZnS core/shell nanotube arrays. *Nanotechnology* 28, 405501. <https://doi.org/10.1088/1361-6528/aa82b0>
- Tereshchenko, A., Bechelany, M., Viter, R., Khranovskyy, V., Smyntyna, V., Starodub, N., Yakimova, R., 2016. Optical biosensors based on ZnO nanostructures: Advantages and perspectives. A review. *Sensors Actuators, B Chem.* 229, 664–677. <https://doi.org/10.1016/j.snb.2016.01.099>
- Tereshchenko, A., Fedorenko, V., Smyntyna, V., Konup, I., Konup, A., Eriksson, M., Yakimova, R., Ramanavicius, A., Balme, S., Bechelany, M., 2017. ZnO films formed by atomic layer deposition as an optical biosensor platform for the detection of Grapevine virus A-type proteins. *Biosens. Bioelectron.* 92, 763–769. <https://doi.org/10.1016/j.bios.2016.09.071>
- Thangaraj, V., Lepoitevin, M., Smietana, M., Balanzat, E., Bechelany, M., Janot, J.M., Vasseur, J.J., Subramanian, S., Balme, S., 2016. Detection of short ssDNA and dsDNA by current-voltage measurements using conical nanopores coated with Al₂O₃ by atomic layer deposition. *Microchim. Acta* 183, 1011–1017. <https://doi.org/10.1007/s00604-015-1706-2>
- Turner, A.P.F., 2013. Biosensors: sense and sensibility. *Chem. Soc. Rev.* 42, 3184. <https://doi.org/10.1039/c3cs35528d>
- Vähä-Nissi, M., Pitkänen, M., Salo, E., Kenttä, E., Tanskanen, A., Sajavaara, T., Putkonen, M., Sievänen, J., Sneck, A., Rättö, M., Karppinen, M., Harlin, A., 2014. Antibacterial and barrier properties of oriented polymer films with ZnO thin films applied with atomic layer deposition at low temperatures. *Thin Solid Films* 562, 331–337. <https://doi.org/10.1016/j.tsf.2014.03.068>
- Van Delft, J.A., Garcia-Alonso, D., Kessels, W.M.M., 2012. Atomic layer deposition for photovoltaics: Applications and prospects for solar cell manufacturing. *Semicond. Sci. Technol.* 27. <https://doi.org/10.1088/0268-1242/27/7/074002>
- Velasco-Garcia, M.N., Mottram, T., 2003. Biosensor Technology addressing Agricultural Problems. *Biosyst. Eng.* 84, 1–12. [https://doi.org/10.1016/S1537-5110\(02\)00236-2](https://doi.org/10.1016/S1537-5110(02)00236-2)

- Vello, T.P., da Silva, L.M.B., Silva, G.O., de Camargo, D.H.S., Corrêa, C.C., Bof Bufon, C.C., 2017. Hybrid organic/inorganic interfaces as reversible label-free platform for direct monitoring of biochemical interactions. *Biosens. Bioelectron.* 87, 209–215. <https://doi.org/10.1016/j.bios.2016.08.050>
- Viter, R., Chaaya, A.A., Iatsunskiy, I., Nowaczyk, G., Kovalevskis, K., Erts, D., Miele, P., Smyntyna, V., Bechelany, M., 2015. Tuning of ZnO 1D nanostructures by atomic layer deposition and electrospinning for optical gas sensor applications. *Nanotechnology* 26, 105501. <https://doi.org/10.1088/0957-4484/26/10/105501>
- Wang, G., Gao, Z., Tang, S., Chen, C., Duan, F., Zhao, S., Lin, S., Feng, Y., Zhou, L., Qin, Y., 2012. Microwave absorption properties of carbon nanocoils coated with highly controlled magnetic materials by atomic layer deposition. *ACS Nano* 6, 11009–11017. <https://doi.org/10.1021/nl304630h>
- Wang, L., Wang, Y., Wong, J.I., Palacios, T., Kong, J., Yang, H.Y., 2014. Functionalized MoS₂ Nanosheet-Based Field-Effect Biosensor for Label-Free Sensitive Detection of Cancer Marker Proteins in Solution. *Small* 10, 1101–1105. <https://doi.org/10.1002/sml.201302081>
- Weber, M., Iatsunskiy, I., Coy, E., Miele, P., Cornu, D., Bechelany, M., 2018. Novel and Facile Route for the Synthesis of Tunable Boron Nitride Nanotubes Combining Atomic Layer Deposition and Annealing Processes for Water Purification. *Adv. Mater. Interfaces* 5, 1800056. <https://doi.org/10.1002/admi.201800056>
- Weber, M., Koonkaew, B., Balme, S., Utke, I., Picaud, F., Iatsunskiy, I., Coy, E., Miele, P., Bechelany, M., 2017. Boron Nitride Nanoporous Membranes with High Surface Charge by Atomic Layer Deposition. *ACS Appl. Mater. Interfaces* 9, 16669–16678. <https://doi.org/10.1021/acsami.7b02883>
- Weber, M.J., MacKus, A.J.M., Verheijen, M.A., Longo, V., Bol, A.A., Kessels, W.M.M., 2014. Atomic layer deposition of high-purity palladium films from Pd(hfac)₂ and H₂ and O₂ plasmas. *J. Phys. Chem. C* 118, 8702–8711. <https://doi.org/10.1021/jp5009412>
- Weber, M.J., Mackus, A.J.M., Verheijen, M.A., van der Marel, C., Kessels, W.M.M., 2012. Supported Core/Shell Bimetallic Nanoparticles Synthesis by Atomic Layer Deposition. *Chem. Mater.* 24, 2973–2977. <https://doi.org/10.1021/cm301206e>
- Weber, M.J., Verheijen, M.A., Bol, A.A., Kessels, W.M.M., 2015. Sub-nanometer dimensions control of core/shell nanoparticles prepared by atomic layer deposition. *Nanotechnology* 26, 94002. <https://doi.org/10.1088/0957-4484/26/9/094002>
- Xiong, S., Kong, L., Huang, J., Chen, X., Wang, Y., 2015. Atomic-layer-deposition-enabled nonwoven membranes with hierarchical ZnO nanostructures for switchable water/oil separations. *J. Memb. Sci.* 493, 478–485. <https://doi.org/10.1016/j.memsci.2015.06.054>
- Yang, P., Tong, X., Wang, G., Gao, Z., Guo, X., Qin, Y., 2015. NiO/SiC nanocomposite prepared by atomic layer deposition used as a novel electrocatalyst for nonenzymatic glucose sensing. *ACS Appl. Mater. Interfaces* 7, 4772–4777. <https://doi.org/10.1021/am508508m>
- Yao, Z., Wang, C., Li, Y., Kim, N.-Y., 2015. AAO-assisted synthesis of highly ordered, large-scale TiO₂ nanowire arrays via sputtering and atomic layer deposition. *Nanoscale Res. Lett.* 10, 166. <https://doi.org/10.1186/s11671-015-0872-9>
- Ye, P.D., Wilk, G.D., Yang, B., Kwo, J., Chu, S.N.G., Nakahara, S., Gossmann, H.-J.L., Mannaerts, J.P., Hong, M., Ng, K.K., Bude, J., 2003. GaAs metal–oxide–semiconductor field-effect transistor with nanometer-thin dielectric grown by atomic layer deposition. *Appl. Phys. Lett.* 83, 180–182.

<https://doi.org/10.1063/1.1590743>

- Yoo, G., Park, H., Kim, M., Song, W.G., Jeong, S., Kim, M.H., Lee, H., Lee, S.W., Hong, Y.K., Lee, M.G., Lee, S., Kim, S., 2017. Real-time electrical detection of epidermal skin MoS₂ biosensor for point-of-care diagnostics. *Nano Res.* 10, 767–775. <https://doi.org/10.1007/s12274-016-1289-1>
- Zafar, S., D’Emic, C., Jagtiani, A., Kratschmer, E., Miao, X., Zhu, Y., Mo, R., Sosa, N., Hamann, H., Shahidi, G., Riel, H., 2018. Silicon Nanowire Field Effect Transistor Sensors with Minimal Sensor-to-Sensor Variations and Enhanced Sensing Characteristics. *ACS Nano* 12, 6577–6587. <https://doi.org/10.1021/acsnano.8b01339>
- Zhang, C., Huang, B., Qian, L., Yuan, S., Wang, S., Chen, R., 2016. Electrochemical Biosensor Based on Nanoporous Au/CoO Core-Shell Material with Synergistic Catalysis. *ChemPhysChem* 17, 98–104. <https://doi.org/10.1002/cphc.201500854>
- Zhang, K., Zhu, Y., Liu, X., Cui, Z., Xianjin, Y., Yeung, K.W.K., Pan, H., Wu, S., 2017. Sr/ZnO doped titania nanotube array: An effective surface system with excellent osteoinductivity and self-antibacterial activity. *Mater. Des.* 130, 403–412. <https://doi.org/10.1016/j.matdes.2017.05.085>
- Zhang, S., Geryak, R., Geldmeier, J., Kim, S., Tsukruk, V. V., 2017. Synthesis, Assembly, and Applications of Hybrid Nanostructures for Biosensing. *Chem. Rev.* 117, 12942–13038. <https://doi.org/10.1021/acs.chemrev.7b00088>
- Zhang, S., Wright, G., Yang, Y., 2000. Materials and techniques for electrochemical biosensor design and construction. *Biosens. Bioelectron.* 15, 273–282. [https://doi.org/10.1016/S0956-5663\(00\)00076-2](https://doi.org/10.1016/S0956-5663(00)00076-2)
- Zhang, X., Zhao, J., Whitney, A. V., Elam, J.W., Van Duyne, R.P., 2006. Ultrastable Substrates for Surface-Enhanced Raman Spectroscopy: Al₂O₃ Overlayers Fabricated by Atomic Layer Deposition Yield Improved Anthrax Biomarker Detection. *J. Am. Chem. Soc.* 128, 10304–10309. <https://doi.org/10.1021/ja0638760>
- Zhao, Y., Pang, F., Dong, Y., Wen, J., Chen, Z., Wang, T., 2013. Refractive index sensitivity enhancement of optical fiber cladding mode by depositing nanofilm via ALD technology. *Opt. Express* 21, 26136. <https://doi.org/10.1364/OE.21.026136>
- Zhao, Y., Zhou, C., Zhang, X., Zhang, P., Dou, Y., Wang, W., Cao, X., Wang, B., Tang, Y., Zhou, S., 2013. Passivation mechanism of thermal atomic layer-deposited Al₂O₃ films on silicon at different annealing temperatures. *Nanoscale Res. Lett.* 8, 114. <https://doi.org/10.1186/1556-276X-8-114>
- Zhu, S., Pang, F., Huang, S., Zou, F., Guo, Q., Wen, J., Wang, T., 2016. High sensitivity refractometer based on TiO₂-Coated adiabatic tapered optical fiber via ALD technology. *Sensors* 16, 1295. <https://doi.org/10.3390/s16081295>