



HAL
open science

Photoacoustic sensing based on resonant mechanical transducers: application to diagnosis in breath

Tarek Seoudi, Diba Ayache, Oifa Benabbad, Julien Charensol, Fanny Pagès, Eric Rosenkrantz, Nicolas Molinari, Michaël Bahriz, Fares Gouzi, Aurore Vicet

► **To cite this version:**

Tarek Seoudi, Diba Ayache, Oifa Benabbad, Julien Charensol, Fanny Pagès, et al.. Photoacoustic sensing based on resonant mechanical transducers: application to diagnosis in breath. ICPPP 2024 - 22st International Conference on Photoacoustic and Photothermal Phenomena, Jul 2024, Coimbra (Portugal), Portugal. <hal-04751842>

HAL Id: hal-04751842

<https://hal.umontpellier.fr/hal-04751842v1>

Submitted on 24 Oct 2024

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Distributed under a Creative Commons CC BY 4.0 - Attribution - International License

Photoacoustic sensing based on resonant mechanical transducers: application to diagnosis in breath

**Seoudi T⁽¹⁾, Ayache D⁽¹⁾, Benabbad O⁽¹⁾, Charensol J⁽¹⁾, Pages F⁽¹⁾, Rosenkantz E⁽¹⁾,
Molinari N⁽³⁾, Bahriz M⁽¹⁾, Gouzi F⁽²⁾, Vicet A^{(1)*}**

(1) IES, Univ. Montpellier, CNRS, Montpellier, France

(2) PhyMedExp INSERM U1046 CNRS UMR 9214, Montpellier University, Department of Clinical Physiology
Montpellier University Hospital, Montpellier France.

(3) IDESP, INSERM, Montpellier University, Montpellier University Hospital, Montpellier, France

*aurore.vicet@umontpellier.fr

Abstract

In this paper we will illustrate our studies on photoacoustic spectroscopy based on mechanical resonators (Quartz tuning forks and MEMS) and their application on breath analysis.

We will describe a several-months measurement campaign at Montpellier Hospital, which focused on diagnosing chronic heart failure using Quartz Enhanced Photoacoustic Spectroscopy (QEPAS).

--

Future medicine needs point of care diagnosis tools dedicated to fast, non-invasive and in situ measurements on patients, potentially at a long distance from any hospital, far from any specialist intervention, to be able to provide adapted treatments and chronic diseases follow up.

Breath analysis might play a substantial role in the future's medicine. It can give access to very simple sample extractions needing neither complex examination nor invasive examination. Breath biomarkers analysis opens the way to various applications from emergency diagnosis to long term follow up of several diseases, from respiratory affections to gastric malabsorptions, intoxications and even cancers [1].

We present here our latest works concerning photoacoustic sensing setups, based on resonant mechanical transducers and their developments on breath sensing. We will describe our last multigas sensing setup based on off-beam QEPAS [2]. QEPAS [3] has acquired a sufficient maturity to reach in situ demonstrations in various fields, such as pollution and environmental sensing [4], or clinical approaches [5].

MEMS photoacoustic spectroscopy is now a real challenger, and progresses were made that position this technique as sensitive as bare QEPAS [6], and its high potential in terms of miniaturisation and integration is a big advantage for future perspectives. We will show our latest developments on silicon-based MEMS designs, fabrication and photoacoustic sensing.

Then, we will present our latest demonstrations on QEPAS to detect several species of medical interest, particularly for chronic heart failure diseases. The setup will be used to investigate four cardiovascular diseases (CVD) biomarkers found in exhaled breath: Carbon Monoxide (CO), Nitric Oxide (NO), isoprene, and acetone, using four different infrared lasers (4.6 μm , 5.2 μm , 11 μm and 8.2 μm). The high resolution and real-time concentration measurements of these biomarkers using QEPAS sensors are combined with the flow rate and capnography provided by a breath sampler used at Montpellier hospital in pulmonary tests. It enables the reconstruction of the complete expirogram for the biomarker and allows tracing their concentration in each compartment of the respiratory system.

Eventually, we will discuss about the observations of this clinical study conducted in Montpellier hospital on both healthy individuals and patients diagnosed with cardiovascular diseases.

Acknowledgments: The authors thank Montpellier University for funding this study via MEDICPAS project (2023-2024) supported by LabEx NUMEV and SENSIR project (2020-2023) supported by I-Site MUSE.

References –

- [1] Kim, K., Jahan, S. A., and Kabir, E. (2012). A review of breath analysis for diagnosis of human health. *Trends in Analytical Chemistry*, 33, 1–8.
- [2] Roman Rousseau, Zeineb Loghmari, Michael Bahriz, Kaim Chamassi, Roland Teissier, Alexei N. Baranov and Aurore Vicet Off-beam QEPAS sensor using a 11 μm DFB-QCL with an optimized acoustic resonator - *Optics Express* 27 (5) 7435-7446, 2019. <https://doi.org/10.1364/OE.27.007435>.
- [3] Kosterev, et al (2002). *Optics letters*, 27(21), 1902-1904.
- [4] Ayache, Diba, et al. "Benzene sensing by quartz enhanced photoacoustic spectroscopy at 14.85 μm ." *Optics Express* 30.4 (2022): 5531-5539.
- [5] N. Maurin, R. Rousseau, W. Trzpił, G. Aoust, M. Hayot, J. Mercier, M. Bahriz, F. Gouzi And A. Vicet First clinical evaluation of a quartz enhanced photo-acoustic CO sensor for human breath analysis. *Sensors and actuators* (319) 128247 (2020). <https://doi.org/10.1016/j.snb.2020.128247>
- [6] Seoudi, T., Charensol, J., Trzpił, W., Pages, F., Ayache, D., Rousseau, R., ... Bahriz, M. (2023). Highly Sensitive Capacitive MEMS for Photoacoustic Gas Trace Detection. *Sensors*, 23(6), 3280. <https://doi.org/10.3390/s23063280>